
Midmark M11 - v1.0.5
Total Cycles: 317

Name: RITTER2
Sterilizer ID: V2313228
08/30/2021 06:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.3 F 0.0
4:00 75.3 F 0.0
6:00 147.7 F 0.2
8:00 161.5 F 0.9
10:00 177.0 F 4.2
12:00 203.8 F 7.8
14:00 209.1 F 11.9
16:00 215.7 F 5.3
18:00 226.9 F 9.2
20:00 237.2 F 13.5
22:00 237.7 F 10.8
24:00 236.0 F 9.2
26:00 250.1 F 15.8
28:00 261.6 F 22.4
30:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.8
2:00 271.6 F 28.9
3:00 272.5 F 30.0
4:00 272.1 F 29.3

Min 271.3 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 217.3 F 0.0
10:00 227.1 F 0.0
15:00 212.7 F 0.0
20:00 201.5 F 0.0
25:00 199.7 F 0.0
30:00 181.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:35
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

08/30/2021 07:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____