

-----  
Midmark M11 - v1.0.5  
Total Cycles: 476

Name: RITTER2  
Sterilizer ID: V2313228  
10/22/2021 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.4 F 0.0  
2:00 160.7 F 2.8  
4:00 184.7 F 6.8  
6:00 206.8 F 12.1  
8:00 225.5 F 7.4  
10:00 238.7 F 12.7  
12:00 237.4 F 10.2  
14:00 235.6 F 8.8  
16:00 249.5 F 15.5  
18:00 261.2 F 22.2  
20:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.9 F 29.6  
2:00 271.1 F 29.0  
3:00 272.6 F 29.8  
4:00 273.4 F 29.9

Min 271.0 F 28.9  
Max 273.5 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 238.2 F 0.0  
10:00 222.6 F 0.0  
15:00 217.7 F 0.0  
20:00 216.0 F 0.0  
25:00 224.6 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:39  
CYCLE COMPLETE

10/22/2021 03:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_