

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2252

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2023 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.8 F 0.0  
2:00 106.4 F 0.0  
4:00 120.6 F 0.2  
6:00 150.4 F 2.6  
8:00 176.0 F 6.0  
10:00 193.3 F 9.8  
12:00 217.5 F 6.9  
14:00 221.9 F 7.6  
16:00 230.8 F 12.1  
18:00 233.2 F 8.8  
20:00 241.1 F 11.4  
22:00 244.4 F 12.9  
24:00 255.9 F 19.0  
26:00 266.3 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.4 F 29.4  
2:00 271.4 F 29.2  
3:00 272.2 F 29.3  
4:00 271.3 F 29.0

Min 271.3 F 28.8  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
5:00 196.6 F 0.0  
10:00 199.6 F 0.0  
15:00 194.0 F 0.0  
20:00 190.1 F 0.0  
25:00 187.5 F 0.0  
30:00 195.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 27:40  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:05:50  
CYCLE COMPLETE

04/14/2023 10:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_