
Midmark M11 - v1.0.5
Total Cycles: 2632

Name: RITTER2
Sterilizer ID: V2313228
07/11/2023 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 179.1 F 2.9
4:00 192.6 F 7.1
6:00 208.4 F 12.0
8:00 224.5 F 7.2
10:00 235.6 F 12.2
12:00 235.2 F 9.8
14:00 237.2 F 9.3
16:00 247.6 F 14.6
18:00 259.4 F 21.1
20:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 271.7 F 29.4
2:00 272.4 F 30.2
3:00 272.7 F 30.3
4:00 272.4 F 29.9

Min 270.8 F 28.8
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 226.1 F 0.0
10:00 219.3 F 0.0
15:00 215.4 F 0.0
20:00 217.9 F 0.0
25:00 213.5 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

07/11/2023 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____