
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001843
8:09 23/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:41
STERILIZING 3:48
136.4 C 222kPa 3:48
136.3 C 223kPa 4:18
136.2 C 221kPa 4:48
136.4 C 221kPa 5:18
136.1 C 221kPa 5:48
136.0 C 221kPa 6:18
136.4 C 224kPa 6:48
136.3 C 223kPa 7:18
136.3 C 223kPa 7:18
VENTING 7:19
AIR DRYING 8:10
CYCLE COMPLETE 8:11

Digital Signature #
BC334220EBE259BB
