
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001957
8:03 15/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:49
STERILIZING 3:52
136.0 C 220kPa 3:52
135.9 C 220kPa 4:23
136.4 C 222kPa 4:53
136.4 C 224kPa 5:23
136.4 C 225kPa 5:53
136.4 C 225kPa 6:23
136.3 C 221kPa 6:53
136.4 C 225kPa 7:23
136.4 C 225kPa 7:23
VENTING 7:23
AIR DRYING 8:12
CYCLE COMPLETE 29:04

Digital Signature #
EE33E98E5CB3EC05
