
Midmark M11 - v1.0.0
Total Cycles: 2463

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2022 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 181.9 F 3.6
4:00 185.4 F 8.4
6:00 208.2 F 10.3
8:00 229.9 F 8.5
10:00 240.0 F 11.6
12:00 241.0 F 11.2
14:00 237.9 F 9.5
16:00 252.4 F 16.4
18:00 264.2 F 23.4
20:00 270.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.5 F 28.9
2:00 272.5 F 29.3
3:00 271.9 F 28.6
4:00 272.8 F 29.3

Min 271.5 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 219.1 F 0.0
10:00 226.6 F 0.0
15:00 230.3 F 0.0
20:00 229.3 F 0.0
25:00 232.0 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

10/20/2022 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____