
Midmark M11 - v1.0.0
Total Cycles: 5224

Name: RITTER 1
Sterilizer ID: V2178198
01/07/2025 02:28 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.7 F 0.0
2:00 168.5 F 3.0
4:00 184.7 F 8.1
6:00 202.5 F 7.6
8:00 226.8 F 8.5
10:00 234.5 F 8.7
12:00 245.7 F 13.5
14:00 244.8 F 12.5
16:00 259.0 F 17.5

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 239F
CHAMBER PRESS 10.0PSI
** ITEMS NOT STERILE **

01/07/2025 02:45 PM