
Midmark M11 - v1.0.5
Total Cycles: 3812

Name: RITTER2
Sterilizer ID: V2313228
07/24/2024 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.0 F 0.0
2:00 93.2 F 0.1
4:00 121.2 F 0.4
6:00 158.2 F 3.0
8:00 178.6 F 6.7
10:00 197.8 F 11.2
12:00 212.7 F 4.1
14:00 224.0 F 8.3
16:00 235.4 F 12.9
18:00 236.5 F 10.0
20:00 236.7 F 8.8
22:00 248.6 F 14.9
24:00 260.5 F 21.6
26:00 271.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.5 F 29.3
2:00 272.3 F 29.5
3:00 272.2 F 29.0
4:00 272.7 F 29.7

Min 271.7 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 204.3 F 0.0
10:00 202.2 F 0.0
15:00 199.5 F 0.0
20:00 197.9 F 0.0
25:00 200.0 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:54
CYCLE COMPLETE

07/24/2024 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____