
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000576
13:27 11/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:21
STERILIZING 4:34
136.0 C 228kPa 4:34
136.4 C 232kPa 5:04
136.4 C 232kPa 5:34
136.4 C 232kPa 6:04
136.3 C 231kPa 6:34
136.4 C 231kPa 7:04
136.2 C 230kPa 7:34
136.2 C 229kPa 8:04
136.2 C 229kPa 8:04
VENTING 8:04
AIR DRYING 8:59
CYCLE COMPLETE 23:59

Digital Signature #
C24E293C652FB49F
