
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
3.1uS / 1.9ppm
CYCLE NUMBER 001872
7:19 19/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:03
STERILIZING 4:11
136.4 C 224kPa 4:11
136.1 C 222kPa 4:41
136.3 C 223kPa 5:11
136.4 C 225kPa 5:41
136.3 C 224kPa 6:11
136.1 C 221kPa 6:41
136.4 C 222kPa 7:11
136.4 C 225kPa 7:41
136.4 C 225kPa 7:41
VENTING 7:42
AIR DRYING 8:31
CYCLE COMPLETE 8:41

Digital Signature #
E0B4E7098493B347
