
Midmark M11 - v1.0.5
Total Cycles: 84

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 199.7 F 5.5
4:00 216.0 F 12.1
6:00 226.9 F 8.6
8:00 236.3 F 9.9
10:00 242.8 F 11.8
12:00 247.9 F 14.5
14:00 261.2 F 21.9
16:00 272.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.2
2:00 271.7 F 29.1
3:00 272.3 F 29.1
4:00 271.6 F 28.9

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 233.0 F 0.0
10:00 221.0 F 0.0
15:00 220.0 F 0.0
20:00 218.7 F 0.0
25:00 219.7 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 16:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:56:02
CYCLE COMPLETE

06/09/2021 01:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____