
Midmark M11 - v1.0.5
Total Cycles: 584

Name: RITTER2
Sterilizer ID: V2313228
11/29/2021 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.9 F 0.0
2:00 160.5 F 3.2
4:00 184.4 F 7.5
6:00 207.2 F 12.9
8:00 219.6 F 7.7
10:00 234.0 F 13.2
12:00 239.0 F 10.7
14:00 236.4 F 9.2
16:00 250.9 F 16.1
18:00 262.4 F 22.9
20:00 273.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 271.8 F 29.8
2:00 272.3 F 29.4
3:00 271.2 F 29.3
4:00 272.7 F 30.5

Min 271.0 F 29.0
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 218.7 F 0.0
10:00 192.2 F 0.0
15:00 191.1 F 0.0
20:00 188.3 F 0.0
25:00 185.0 F 0.0
30:00 182.2 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

11/30/2021 07:37 AM
ACCEPTED: X
REJECTED:
Operator: U02