
Midmark M11 - v1.0.0
Total Cycles: 4151

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2023 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 74.4 F 0.2
4:00 90.8 F 1.4
6:00 123.9 F 3.8
8:00 155.5 F 6.9
10:00 177.1 F 10.4
12:00 205.2 F 4.9
14:00 210.8 F 4.8
16:00 222.5 F 8.2
18:00 231.4 F 12.0
20:00 232.0 F 7.3
22:00 243.3 F 12.3
24:00 239.7 F 9.9
26:00 252.2 F 16.0
28:00 262.9 F 22.2
30:00 272.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.0
1:00 272.3 F 28.2
2:00 272.4 F 28.7
3:00 271.0 F 27.9
4:00 272.2 F 28.4

Min 271.0 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.4 F 0.0
5:00 231.1 F 0.0
10:00 223.9 F 0.0
15:00 191.2 F 0.0
20:00 189.7 F 0.0
25:00 193.8 F 0.0
30:00 197.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

12/22/2023 10:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____