
Midmark M11 - v1.0.5
Total Cycles: 1062

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 179.0 F 3.5
4:00 199.3 F 8.8
6:00 223.0 F 8.9
8:00 235.0 F 10.5
10:00 233.3 F 8.2
12:00 240.0 F 10.5
14:00 247.0 F 14.0
16:00 259.2 F 20.8
18:00 270.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 29.4
2:00 272.7 F 29.9
3:00 271.3 F 29.2
4:00 272.7 F 29.7

Min 271.3 F 29.1
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 236.2 F 0.0
10:00 223.7 F 0.0
15:00 216.2 F 0.0
20:00 185.7 F 0.0
25:00 180.2 F 0.0
30:00 177.5 F 0.0

Drying Complete

FILLING: 1:46
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/03/2022 11:11 AM
ACCEPTED: X
REJECTED:
Operator: U02