
Midmark M11 - v1.0.5
Total Cycles: 2776

Name: RITTER2
Sterilizer ID: V2313228
08/16/2023 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 164.2 F 0.0
4:00 175.5 F 0.0
6:00 197.8 F 0.0
8:00 212.2 F 2.7
10:00 223.4 F 6.6
12:00 234.0 F 10.9
14:00 233.3 F 8.1
16:00 244.4 F 13.4
18:00 243.4 F 12.3
20:00 242.5 F 11.6
22:00 254.9 F 18.0
24:00 265.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.7 F 29.2
2:00 271.6 F 28.6
3:00 272.8 F 29.2
4:00 271.8 F 28.5

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 211.0 F 0.0
10:00 213.1 F 0.0
15:00 210.9 F 0.0
20:00 207.3 F 0.0
25:00 204.5 F 0.0
30:00 202.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 25:48
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

08/16/2023 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____