
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002013
14:10 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:45
136.2 C 222kPa 3:45
136.4 C 225kPa 4:15
136.4 C 225kPa 4:45
136.0 C 222kPa 5:15
136.3 C 224kPa 5:45
136.4 C 224kPa 6:15
136.4 C 225kPa 6:45
136.4 C 224kPa 7:15
136.4 C 224kPa 7:15
VENTING 7:15
AIR DRYING 8:06
CYCLE COMPLETE 35:53

Digital Signature #
5E975D86C6A6144D
