
Midmark M11 - v1.0.0
Total Cycles: 4154

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 157.9 F 2.8
4:00 179.3 F 6.3
6:00 199.2 F 10.3
8:00 223.1 F 5.3
10:00 229.9 F 8.5
12:00 237.3 F 13.2
14:00 238.7 F 10.0
16:00 236.5 F 8.8
18:00 248.4 F 14.1
20:00 259.9 F 20.5
22:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.7 F 28.4
2:00 272.1 F 28.6
3:00 272.3 F 28.9
4:00 271.4 F 28.6

Min 271.1 F 28.3
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 233.9 F 0.0
10:00 219.8 F 0.0
15:00 228.8 F 0.0
20:00 215.2 F 0.0
25:00 224.4 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:29
CYCLE COMPLETE

12/22/2023 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____