
Midmark M11 - v1.0.0
Total Cycles: 1043

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2021 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 169.7 F 1.8
4:00 179.2 F 5.3
6:00 200.3 F 9.9
8:00 219.0 F 6.8
10:00 229.6 F 9.4
12:00 236.1 F 9.6
14:00 241.3 F 11.8
16:00 239.4 F 10.2
18:00 252.9 F 16.9
20:00 264.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.4 F 28.3
3:00 272.2 F 29.2
4:00 272.1 F 28.8

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 208.6 F 0.0
10:00 214.6 F 0.0
15:00 216.1 F 0.0
20:00 220.2 F 0.0
25:00 221.8 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

07/22/2021 09:23 AM
ACCEPTED: X
REJECTED:
Operator: U02