
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002622
12:58 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:24
136.0 C 221kPa 3:24
136.0 C 221kPa 3:55
136.2 C 223kPa 4:25
136.2 C 223kPa 4:55
136.4 C 225kPa 5:25
136.1 C 222kPa 5:55
136.4 C 224kPa 6:25
136.3 C 225kPa 6:55
136.3 C 225kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 15:17

Digital Signature #
E3BF7115972AB71B
