
Midmark M11 - v1.0.5
Total Cycles: 1432

Name: RITTER2
Sterilizer ID: V2313228
08/30/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 188.0 F 2.9
4:00 203.3 F 8.0
6:00 221.0 F 13.4
8:00 230.2 F 8.9
10:00 237.6 F 10.4
12:00 243.3 F 12.9
14:00 243.1 F 12.3
16:00 256.1 F 19.2
18:00 267.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.1
1:00 272.7 F 29.8
2:00 271.6 F 29.1
3:00 272.4 F 29.6
4:00 271.6 F 29.2

Min 271.6 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 231.7 F 0.0
10:00 229.5 F 0.0
15:00 230.0 F 0.0
20:00 225.6 F 0.0
25:00 220.9 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

08/30/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____