
Midmark M11 - v1.0.0
Total Cycles: 1386

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2021 01:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 178.0 F 3.6
4:00 198.7 F 7.8
6:00 221.2 F 12.2
8:00 230.8 F 8.6
10:00 239.3 F 11.0
12:00 241.6 F 11.7
14:00 239.6 F 10.2
16:00 253.7 F 17.2
18:00 265.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.3 F 28.8
2:00 271.3 F 28.3
3:00 272.4 F 29.0
4:00 271.6 F 28.3

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 253.7 F 0.0
10:00 232.3 F 0.0
15:00 225.8 F 0.0
20:00 222.9 F 0.0
25:00 223.3 F 0.0
Dry Cycle Aborted

FILLING: 2:15
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:08
DRYING: 29:51
TOTAL CYCLE: 00:58:54

11/18/2021 02:46 PM
ACCEPTED: X
REJECTED:
Operator: U02