
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002296
12:05 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:54
STERILIZING 2:58
136.4 C 224kPa 2:58
136.3 C 225kPa 3:28
136.3 C 225kPa 3:58
136.4 C 226kPa 4:28
136.4 C 225kPa 4:58
136.4 C 226kPa 5:28
136.4 C 223kPa 5:58
136.4 C 226kPa 6:28
136.4 C 226kPa 6:28
VENTING 6:28
AIR DRYING 7:19
CYCLE COMPLETE 37:19

Digital Signature #
BFFE5CE52F8A65DD
