
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001476
12:54 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:43
STERILIZING 3:45
136.0 C 223kPa 3:45
136.2 C 224kPa 4:15
136.2 C 223kPa 4:45
136.2 C 225kPa 5:15
136.4 C 226kPa 5:45
136.4 C 227kPa 6:15
136.4 C 226kPa 6:45
136.4 C 227kPa 7:15
136.4 C 227kPa 7:15
VENTING 7:15
AIR DRYING 8:09
CYCLE COMPLETE 8:20

Digital Signature #
36CF9B0B42F36A0E
