

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002282  
12:33 18/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:25  
STERILIZING 3:27  
136.0 C 221kPa 3:27  
136.2 C 224kPa 3:57  
136.4 C 226kPa 4:27  
136.3 C 225kPa 4:57  
136.3 C 225kPa 5:27  
136.4 C 227kPa 5:57  
136.4 C 226kPa 6:27  
136.3 C 225kPa 6:57  
136.3 C 225kPa 6:57  
VENTING 6:58  
AIR DRYING 7:51  
CYCLE COMPLETE 17:53  
-----

Digital Signature #  
FD033439A603FA2D  
-----