

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4555

Name: RITTER 1  
Sterilizer ID: V2178198  
05/15/2024 09:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
2:00 174.6 F 2.5  
4:00 174.6 F 6.1  
6:00 193.1 F 10.0  
8:00 218.0 F 7.3  
10:00 216.9 F 7.1  
12:00 228.9 F 11.4  
14:00 229.9 F 7.5  
16:00 243.4 F 13.0  
18:00 241.3 F 11.3  
20:00 253.3 F 17.2  
22:00 263.8 F 23.3  
24:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.4 F 29.2  
2:00 271.7 F 28.6  
3:00 271.2 F 28.6  
4:00 272.3 F 29.2

Min 271.1 F 28.3  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 215.6 F 0.0  
10:00 215.2 F 0.0  
15:00 225.7 F 0.0  
20:00 231.6 F 0.0  
25:00 227.4 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:03:20  
CYCLE COMPLETE

05/15/2024 10:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_