

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5249

Name: RITTER 1  
Sterilizer ID: V2178198  
01/16/2025 11:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.0 F 0.0  
2:00 127.3 F 0.0  
4:00 140.4 F 0.0  
6:00 171.1 F 1.8  
8:00 194.0 F 5.4  
10:00 207.9 F 9.1  
12:00 224.3 F 13.2  
14:00 226.5 F 8.0  
16:00 237.6 F 12.8  
18:00 236.7 F 9.9  
20:00 237.1 F 8.9  
22:00 247.3 F 14.0  
24:00 258.7 F 20.2  
26:00 269.1 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.3 F 28.9  
2:00 271.5 F 28.6  
3:00 272.3 F 28.9  
4:00 271.4 F 28.4

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
5:00 197.9 F 0.0  
10:00 194.4 F 0.0  
15:00 189.9 F 0.0  
20:00 202.4 F 0.0  
25:00 211.1 F 0.0  
30:00 216.5 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 27:06  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:05:38  
CYCLE COMPLETE

01/16/2025 12:46 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_