
Midmark M11 - v1.0.0
Total Cycles: 1426

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2021 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.3 F 0.0
2:00 143.1 F 1.3
4:00 161.5 F 4.1
6:00 190.4 F 8.1
8:00 211.7 F 13.1
10:00 224.0 F 7.1
12:00 239.0 F 12.1
14:00 235.2 F 9.2
16:00 237.5 F 9.4
18:00 247.5 F 14.1
20:00 259.5 F 20.8
22:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.6 F 28.8
3:00 272.6 F 29.2
4:00 271.4 F 28.7

Min 271.4 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 235.3 F 0.0
10:00 226.8 F 0.0
15:00 227.6 F 0.0
20:00 210.7 F 0.0
25:00 200.5 F 0.0
30:00 193.4 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

12/06/2021 03:22 PM
ACCEPTED: X
REJECTED:
Operator: U02