
Midmark M11 - v1.0.0
Total Cycles: 3284

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 189.0 F 1.8
4:00 202.8 F 6.7
6:00 220.0 F 12.3
8:00 226.8 F 8.0
10:00 240.4 F 13.4
12:00 241.0 F 11.3
14:00 239.3 F 10.0
16:00 253.4 F 16.9
18:00 265.1 F 23.8
20:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 28.8
2:00 272.1 F 29.0
3:00 272.3 F 28.7
4:00 272.4 F 29.1

Min 271.6 F 28.4
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 235.3 F 0.0
10:00 228.8 F 0.0
15:00 226.6 F 0.0
20:00 225.4 F 0.0
25:00 231.2 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

05/25/2023 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____