
Midmark M11 - v1.0.5
Total Cycles: 287

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.5 F 0.0
2:00 144.9 F 1.6
4:00 159.8 F 4.9
6:00 185.9 F 9.7
8:00 221.3 F 5.6
10:00 227.1 F 9.2
12:00 236.0 F 9.9
14:00 241.5 F 11.9
16:00 239.9 F 10.7
18:00 253.4 F 17.4
20:00 265.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 271.8 F 28.7
2:00 272.2 F 29.0
3:00 272.0 F 29.4
4:00 271.2 F 29.1

Min 271.0 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
5:00 229.6 F 0.0
10:00 215.8 F 0.0
15:00 212.4 F 0.0
20:00 213.6 F 0.0
25:00 218.3 F 0.0
Dry Cycle Aborted

FILLING: 1:33
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:03
DRYING: 27:05
TOTAL CYCLE: 00:57:22

08/19/2021 09:10 AM
ACCEPTED: X
REJECTED:
Operator: U02