
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000672
10:33 21/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:42
136.5 C 221kPa 4:42
136.4 C 223kPa 5:12
136.3 C 224kPa 5:42
136.2 C 221kPa 6:12
136.2 C 223kPa 6:42
136.2 C 220kPa 7:12
136.1 C 220kPa 7:42
136.4 C 224kPa 8:12
136.4 C 224kPa 8:12
VENTING 8:12
AIR DRYING 9:07
CYCLE COMPLETE 9:14

Digital Signature #
B6C78B7D319521F7
