

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000555  
15:08 23/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:08  
STERILIZING 4:19  
136.4 C 232kPa 4:19  
136.4 C 233kPa 4:49  
136.2 C 229kPa 5:19  
136.2 C 232kPa 5:49  
136.4 C 233kPa 6:19  
136.4 C 233kPa 6:49  
136.4 C 233kPa 7:19  
136.3 C 232kPa 7:49  
136.3 C 232kPa 7:49  
VENTING 7:49  
AIR DRYING 8:43  
CYCLE COMPLETE 23:43  
-----

Digital Signature #  
EE20629D6D8BCFC1  
-----