
Midmark M11 - v1.0.0
Total Cycles: 5478

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2025 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.2 F 0.0
2:00 179.1 F 2.3
4:00 185.9 F 6.0
6:00 194.9 F 9.9
8:00 218.7 F 8.2
10:00 219.4 F 6.0
12:00 230.7 F 10.2
14:00 236.4 F 8.6
16:00 242.4 F 11.5
18:00 238.8 F 9.6
20:00 251.2 F 15.6
22:00 261.6 F 21.6
24:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 271.5 F 28.5
2:00 272.8 F 29.2
3:00 272.8 F 28.8
4:00 271.6 F 28.4

Min 270.9 F 27.9
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 249.0 F 0.0
10:00 227.3 F 0.0
15:00 212.9 F 0.0
20:00 208.0 F 0.0
25:00 211.1 F 0.0
30:00 205.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

04/04/2025 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____