
Midmark M11 - v1.0.5
Total Cycles: 483

Name: RITTER2
Sterilizer ID: V2313228
10/26/2021 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 149.6 F 1.6
4:00 155.3 F 4.7
6:00 176.6 F 9.2
8:00 208.9 F 6.8
10:00 216.3 F 9.5
12:00 228.4 F 8.6
14:00 242.5 F 13.1
16:00 242.6 F 12.0
18:00 255.4 F 18.7
20:00 266.7 F 25.8
22:00 273.4 F 30.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.1
1:00 272.5 F 30.7
2:00 272.0 F 30.0
3:00 273.4 F 30.9
4:00 272.2 F 30.5

Min 271.7 F 29.9
Max 274.2 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 232.6 F 0.0
10:00 222.6 F 0.0
15:00 222.4 F 0.0
20:00 224.9 F 0.0
25:00 231.6 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

10/26/2021 10:50 AM
ACCEPTED: X
REJECTED:
Operator: U01