
Midmark M11 - v1.0.0
Total Cycles: 4540

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2024 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 71.2 F 0.0
4:00 85.9 F 0.2
6:00 123.7 F 1.9
8:00 160.5 F 4.6
10:00 180.4 F 7.9
12:00 203.5 F 11.9
14:00 215.9 F 3.7
16:00 227.4 F 7.5
18:00 236.4 F 11.3
20:00 228.4 F 6.2
22:00 241.8 F 11.7
24:00 237.5 F 9.1
26:00 250.5 F 15.4
28:00 261.4 F 21.6
30:00 271.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.7 F 28.4
2:00 271.9 F 28.4
3:00 272.4 F 28.5
4:00 272.5 F 28.8

Min 271.5 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 222.3 F 0.0
10:00 201.6 F 0.0
15:00 195.2 F 0.0
20:00 181.6 F 0.0
25:00 180.8 F 0.0
30:00 186.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:13
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

05/10/2024 09:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____