
Midmark M11 - v1.0.0
Total Cycles: 4285

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2024 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.5 F 0.0
2:00 98.3 F 0.6
4:00 104.9 F 2.4
6:00 142.7 F 4.7
8:00 157.5 F 7.7
10:00 183.5 F 11.2
12:00 216.1 F 3.9
14:00 215.1 F 6.2
16:00 227.4 F 9.9
18:00 236.6 F 10.2
20:00 237.0 F 10.3
22:00 237.6 F 8.7
24:00 246.4 F 13.5
26:00 256.7 F 19.1
28:00 266.1 F 24.9
30:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.3 F 28.9
2:00 272.8 F 29.2
3:00 272.5 F 29.4
4:00 271.6 F 28.9

Min 271.6 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 233.1 F 0.0
10:00 234.8 F 0.0
15:00 219.2 F 0.0
20:00 228.9 F 0.0
25:00 230.2 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

02/08/2024 02:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____