
Midmark M11 - v1.0.5
Total Cycles: 2406

Name: RITTER2
Sterilizer ID: V2313228
05/24/2023 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.1 F 0.0
2:00 165.3 F 3.2
4:00 174.0 F 6.7
6:00 197.3 F 11.7
8:00 215.9 F 5.7
10:00 228.6 F 10.6
12:00 231.2 F 7.7
14:00 243.0 F 12.9
16:00 244.1 F 12.8
18:00 256.5 F 19.4
20:00 267.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 273.0 F 29.9
2:00 271.8 F 29.0
3:00 272.7 F 30.0
4:00 272.5 F 29.5

Min 271.6 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 215.7 F 0.0
10:00 227.9 F 0.0
15:00 229.9 F 0.0
20:00 229.7 F 0.0
25:00 230.2 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

05/24/2023 10:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____