

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4365

Name: RITTER2  
Sterilizer ID: V2313228  
02/20/2025 04:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.6 F 0.0  
2:00 106.1 F 0.7  
4:00 118.3 F 2.4  
6:00 148.5 F 5.3  
8:00 175.9 F 8.4  
10:00 193.3 F 11.8  
12:00 213.7 F 4.5  
14:00 219.2 F 8.8  
16:00 232.6 F 13.2  
18:00 234.4 F 9.5  
20:00 240.7 F 10.9  
22:00 243.4 F 12.7  
24:00 254.0 F 18.3  
26:00 264.0 F 24.2  
28:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.3 F 30.3  
2:00 271.6 F 30.0  
3:00 271.8 F 29.7  
4:00 272.6 F 30.0

Min 271.0 F 29.1  
Max 273.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 200.3 F 0.0  
10:00 202.9 F 0.0  
15:00 212.6 F 0.0  
20:00 208.5 F 0.0  
25:00 203.7 F 0.0  
30:00 208.7 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 28:10  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:07:06  
CYCLE COMPLETE

02/20/2025 05:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_