
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000777
14:50 09/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:43
136.4 C 223kPa 4:43
136.4 C 224kPa 5:13
136.4 C 218kPa 5:43
136.4 C 223kPa 6:13
136.3 C 223kPa 6:43
136.4 C 225kPa 7:13
136.2 C 223kPa 7:43
136.2 C 223kPa 8:13
136.2 C 223kPa 8:13
VENTING 8:13
AIR DRYING 9:10
CYCLE COMPLETE 9:10

Digital Signature #
EFA4B7DF61A52EE5
