
Midmark M11 - v1.0.0
Total Cycles: 1840

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2022 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 200.3 F 4.7
4:00 205.7 F 10.0
6:00 218.4 F 5.7
8:00 230.0 F 10.5
10:00 231.5 F 7.2
12:00 243.0 F 12.4
14:00 243.9 F 12.6
16:00 256.8 F 19.5
18:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.4
2:00 271.4 F 28.7
3:00 272.5 F 29.5
4:00 271.9 F 28.9

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 207.2 F 0.0
10:00 234.1 F 0.0
15:00 233.7 F 0.0
20:00 228.9 F 0.0
25:00 227.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

04/19/2022 11:15 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____