

-----  
Midmark M11 - v1.0.5  
Total Cycles: 731

Name: RITTER2  
Sterilizer ID: V2313228  
01/26/2022 11:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
2:00 185.7 F 4.1  
4:00 206.4 F 9.8  
6:00 225.0 F 6.7  
8:00 239.5 F 12.1  
10:00 239.3 F 10.4  
12:00 237.8 F 9.5  
14:00 253.0 F 16.9  
16:00 265.4 F 24.3  
18:00 272.5 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.9 F 29.3  
2:00 272.7 F 29.7  
3:00 271.8 F 28.9  
4:00 272.7 F 29.5  
  
Min 271.7 F 28.9  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
5:00 227.9 F 0.0  
10:00 225.0 F 0.0  
15:00 190.2 F 0.0  
Dry Cycle Aborted

FILLING: 2:22  
HEATING: 18:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 16:23  
TOTAL CYCLE: 00:44:19

01/26/2022 12:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_