
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000681
7:06 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:04
STERILIZING 3:10
136.4 C 231kPa 3:10
136.2 C 229kPa 3:40
136.4 C 232kPa 4:10
136.3 C 229kPa 4:40
136.4 C 229kPa 5:10
136.4 C 231kPa 5:40
136.2 C 231kPa 6:10
136.2 C 229kPa 6:40
136.2 C 229kPa 6:40
VENTING 6:40
AIR DRYING 7:27
CYCLE COMPLETE 7:30

Digital Signature #
388B692236B4E65D
