
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000854
9:01 23/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:46
136.5 C 223kPa 4:46
136.4 C 225kPa 5:16
136.4 C 224kPa 5:46
136.4 C 222kPa 6:16
136.4 C 223kPa 6:46
136.2 C 224kPa 7:16
136.4 C 221kPa 7:46
136.4 C 219kPa 8:16
136.4 C 219kPa 8:16
VENTING 8:16
AIR DRYING 9:13
CYCLE COMPLETE 9:14

Digital Signature #
0AEE5C11B4E0F79F
