
Midmark M11 - v1.0.5
Total Cycles: 4235

Name: RITTER2
Sterilizer ID: V2313228
12/20/2024 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 160.9 F 2.5
4:00 183.7 F 6.4
6:00 200.3 F 10.7
8:00 214.2 F 5.1
10:00 226.0 F 8.5
12:00 237.7 F 13.2
14:00 238.9 F 10.8
16:00 237.3 F 9.5
18:00 250.7 F 15.9
20:00 261.8 F 22.3
22:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.4
2:00 271.7 F 28.7
3:00 272.6 F 29.6
4:00 272.3 F 29.0

Min 271.6 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 219.6 F 0.0
10:00 214.9 F 0.0
15:00 210.9 F 0.0
20:00 210.8 F 0.0
25:00 213.7 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

12/20/2024 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____