
Midmark M11 - v1.0.0
Total Cycles: 5177

Name: RITTER 1
Sterilizer ID: V2178198
12/11/2024 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 77.7 F 0.0
4:00 89.3 F 0.0
6:00 124.5 F 0.0
8:00 158.8 F 0.1
10:00 185.9 F 0.3
12:00 203.5 F 0.7
14:00 216.9 F 4.0
16:00 224.7 F 6.8
18:00 233.6 F 10.3
20:00 235.9 F 9.2
22:00 241.8 F 11.7
24:00 238.5 F 9.7
26:00 237.0 F 8.5
28:00 250.3 F 15.2
30:00 262.4 F 22.2
32:00 272.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.8 F 28.8
2:00 271.7 F 28.0
3:00 272.8 F 29.0
4:00 272.3 F 28.3

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 164.9 F 0.0
5:00 190.2 F 0.0
10:00 196.2 F 0.0
15:00 193.5 F 0.0
20:00 192.0 F 0.0
25:00 204.4 F 0.0
30:00 214.6 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 32:18
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

12/11/2024 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____