
Midmark M11 - v1.0.0
Total Cycles: 5381

Name: RITTER 1
Sterilizer ID: V2178198
03/03/2025 10:43 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.8 F 0.0
2:00 70.8 F 0.3
4:00 93.0 F 1.7
6:00 130.2 F 5.2
8:00 158.6 F 8.8
10:00 182.4 F 13.2
12:00 203.5 F 3.5
14:00 216.9 F 7.2
16:00 228.6 F 11.2
18:00 226.7 F 5.8
20:00 240.2 F 11.5
22:00 236.4 F 8.7
24:00 250.8 F 15.6
26:00 263.1 F 22.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.9 F 28.9
2:00 272.1 F 28.6
3:00 272.1 F 28.2

Min 271.6 F 28.0
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 173.6 F 0.0
10:00 170.0 F 0.0
15:00 172.0 F 0.0
20:00 178.4 F 0.0
25:00 183.2 F 0.0
30:00 186.6 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 27:56
STERILIZING: 3:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

03/03/2025 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____