
Midmark M11 - v1.0.5
Total Cycles: 985

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.9 F 0.0
2:00 143.5 F 0.1
4:00 168.5 F 1.0
6:00 194.3 F 5.5
8:00 210.6 F 10.1
10:00 228.0 F 7.1
12:00 238.4 F 11.5
14:00 236.6 F 9.4
16:00 238.3 F 9.5
18:00 250.0 F 15.5
20:00 262.3 F 22.5
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.7 F 29.6
2:00 272.6 F 29.7
3:00 271.8 F 29.4
4:00 272.1 F 29.1

Min 271.0 F 28.5
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 221.5 F 0.0
10:00 239.7 F 0.0
15:00 224.1 F 0.0
20:00 228.8 F 0.0
25:00 229.9 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

04/13/2022 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____