

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3130

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2023 05:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.5 F 0.0  
2:00 146.1 F 1.7  
4:00 162.8 F 5.1  
6:00 188.0 F 9.4  
8:00 206.4 F 9.4  
10:00 220.4 F 7.9  
12:00 232.2 F 12.6  
14:00 234.1 F 9.8  
16:00 236.5 F 9.3  
18:00 247.3 F 14.6  
20:00 259.1 F 21.1  
22:00 269.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.6 F 29.9  
2:00 271.5 F 29.2  
3:00 271.9 F 29.9  
4:00 272.6 F 29.8

Min 271.2 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
5:00 206.8 F 0.0  
10:00 212.0 F 0.0  
15:00 212.6 F 0.0  
20:00 209.3 F 0.0  
25:00 207.1 F 0.0  
30:00 203.2 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 22:58  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:01:44  
CYCLE COMPLETE

11/15/2023 06:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_