

-----  
Midmark M11 - v1.0.5  
Total Cycles: 377

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2021 08:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.7 F 0.0  
2:00 145.3 F 1.6  
4:00 149.1 F 4.8  
6:00 177.4 F 9.5  
8:00 200.8 F 5.2  
10:00 217.3 F 9.8  
12:00 228.7 F 8.0  
14:00 240.9 F 12.4  
16:00 244.7 F 12.7  
18:00 257.3 F 19.7  
20:00 268.2 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 271.8 F 30.1  
2:00 271.5 F 29.6  
3:00 272.0 F 30.0  
4:00 272.1 F 30.3  
  
Min 270.4 F 28.9  
Max 273.6 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 239.2 F 0.0  
10:00 225.7 F 0.0  
15:00 228.1 F 0.0  
20:00 230.4 F 0.0  
25:00 196.4 F 0.0  
30:00 197.6 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:60  
STERILIZING: 4:00  
VENTING: 2:53  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

09/20/2021 09:06 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01