

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3039

Name: RITTER2  
Sterilizer ID: V2313228  
10/20/2023 04:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
2:00 188.4 F 4.2  
4:00 210.1 F 10.4  
6:00 224.3 F 6.4  
8:00 239.1 F 12.9  
10:00 240.6 F 11.5  
12:00 240.7 F 11.1  
14:00 255.6 F 18.7  
16:00 268.1 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.4 F 29.1  
2:00 272.4 F 29.6  
3:00 272.1 F 28.9  
4:00 272.5 F 29.6  
  
Min 271.6 F 28.6  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 230.4 F 0.0  
10:00 214.2 F 0.0  
15:00 228.8 F 0.0  
20:00 228.9 F 0.0  
25:00 230.1 F 0.0  
30:00 228.6 F 0.0  
  
Drying Complete

FILLING: 2:41  
HEATING: 17:09  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:56:56  
CYCLE COMPLETE

10/20/2023 04:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_